



Terms of Reference for the Outsourcing of **Endoscopy Machine**

1. SPECIFICATIONS:

- With HD Video Processor and CLV
- Light Source with Pigtail, Keyboard, Light Control Cable
- Water Bottle, RGB Cable and Power Cords
- With Video Colonoscope with 12.8mm Outer Diameter
- 3.7mm Working Channel
- 170 degree angle of View, 168cm Working Length
- Angulation 180°/180° Up/Down and 160°/160° Right/Left
- With Video Gastroscope with 8.8mm Outer Diameter
- 2.8mm Working Channel
- 140° Angle of View, 103cm Working Length, Angulation 210°/90° Up/Down and 100°/100° Right Left
- With medical grade monitor with stand
- With procedure cart
- Biopsy forcep 230cm in length
- Cleaning brush
- Injector force max injection needles
- Grasping forceps/ Shark tooth
- Hemoclip
- Deployer

2. MAINTENANCE AND SERVICE

- The winning provider shall ensure the availability of qualified technician and trained personnel to maintain the equipment.
- Shall submit the Schedule of corrective maintenance upon facilitation of agreement.
- The supplier shall provide one unit and shall provide additional unit for back-up if needed.
- The company's technicians shall be on-call 24/7 including holidays if the end-user needs assistance.
- The company shall repair the machine or provide a replacement unit not later than twenty-four (24) hours from the time the hospital report the issues.

3. OTHER PROVISION

- The proposal shall be evaluated by the end-user for consideration.