


**HOSPITAL BIDS AND AWARDS COMMITTEE**  
**SUPPLEMENTAL BID BULLETIN NO. 1**  
**IB 2021-007**  
**PRE-BID CONFERENCE (MAY 14, 2021)**

**I. GENERAL INSTRUCTIONS**

- Bidders are required to submit one (1) soft copy of the original documents clearly scanned in multiple pages in one (1) PDF File per the three required files to the designated Google Form link.
- Documentary requirements on the electronic bidding is the same as in the manual bidding as provided in the 2016 Revised Implementing Rules and Regulations (IRR) of Republic Act 9184 (RA 9184), otherwise known as the “Government Procurement Reform Act” and Resolution No. 09-2020 of the Government Procurement Policy Board entitled “Approving Measures for the Efficient Conduct of Procurement Activities During a State of Calamity, or Implementation of Community Quarantine or Similar Restrictions”
- The BAC is using the 6<sup>th</sup> Edition of the Philippine Bidding Documents and all Amendments Forms therein.
- Due to limited slots in Zoom, only one (1) representative per bidder will be allowed to enter.
- Bidder shall only provide the password of their encrypted file upon being called and acknowledged in the Online Bidding Conference by the BAC before the opening of said Bidder’s bid documents. Bidder’s video should be on during this process.
- For the issuance of the Official Receipt, bidders may get the same at the Cash Collection Section or Procurement Section, during office hours, M-F; 8:00AM-5:00PM, except holidays, and upon showing of proof of payment. For the purpose of the bidding, those who submitted the proof of payment early will be verified by our Cash Collection Section and will be provided a soft copy of the Official Receipt, otherwise, other proof of payment will be acceptable subject to later verification by the Cash Collection Section.
- The winning bidder is required to produce two (2) hard copies of the scanned document anytime before the award. The same can be personally sent or via courier. Any discrepancy on the submitted soft copy as bidder during the bidding conference and hard copies as winning bidder is a ground for disqualification.
- The scanned copy should have the required original signature. E-signature is not allowed.
- Bidders must submit updated Tax clearance per E.O. No. 398, s. 2005, as finally reviewed and approved by the Bureau of Internal Revenue (BIR).
- It is understood that the Contractor is legally responsible to deliver all issued purchase order/s and failure to deliver the first Purchase Order as scheduled shall mean automatic cancellation of the PO and Notice of Award (NOA). Upon cancellation, the BAC shall proceed to qualify the second lowest bidder if applicable; or proceed to Negotiated Procurement. The Winning Contractor who failed to deliver shall shoulder the price difference (from the second lowest bidder) of the item in addition to the acquired liquidated damages.
- In lieu of the Certificate of Ongoing Projects/Accomplishments, Bidders must present their progress report or percentage of partial accomplishment, together with their contracts, awards, and proof of deliveries such as sales invoice.





- Bidders must specify in the Technical Specifications Form under Bidder’s Offer, Technical Specifications Column the exact specifications they are offering and must indicate the reference page number in the user manual submitted against each of the individual parameters. Whereas, bidders must state either “Comply” or “Not Comply” under the Bidder’s Offer, Statement of Compliance Column. The bids with incomplete and incorrect filled out Technical Specifications Form will be disqualified.
- Bidder must have completed, within the period specified in the Invitation to Bid, an SLCC that is similar to the contract to be bid, and whose value, adjusted to current prices using the Philippine Statistics Authority (PSA) consumer price indices, must be at least fifty percent (50%) of the ABC. However, in the case of Expendable Supplies, said SLCC must be at least twenty five percent (25%) of the ABC. Service will fall on the Expendable Supplies.
- If there is no single completed contract for the required amount, the bidder may submit at least two (2) completed similar contracts and the aggregate contract amounts should be equivalent to at least the percentage of the ABC as required above and the largest of these similar contracts must be equivalent to at least half of the percentage of the ABC as required.
- A contract shall only be considered “similar” to the contract to be bid if it is of similar nature.
- Contracts similar to the Project should be completed within two (2) years to the deadline for the submission and receipt of bids.

II. Amendment on Specifications as per Pre-Bidding Conference conducted last May 14, 2021 as follows:

**A. IB2021-007A SUPPLY AND DELIVERY OF 1 UNIT ANESTHESIA MACHINE-REBID**

<b>FROM</b>	<b>TO</b>
<b>ANESTHESIA MACHINE</b>  4. BREATHING SYSTEM: 4.5. Pneumatically driven	<b>ANESTHESIA MACHINE</b>  4. BREATHING SYSTEM 4.5 Pneumatically driven <b>or electrical driven but all pneumatic functions continue to be available (APL valve, breathing pressure gauge, cylinder and pipeline gauges, fresh gas and agent delivery, and O2, AIR, flow meters) in case of power failure and even if the backup battery is completely discharged.</b>
6. VAPORIZER  6.1 Both for Sevoflurane and Isoflurane	6. VAPORIZER  6.1 Both for Sevoflurane and Isoflurane
10. WRITING TABLE/DRAWERS/MOUNTING  10.2. With a retractable writing table that can hold at least 10kgs.	10. WRITING TABLE/DRAWERS/MOUNTING  10. 2. With a <b>fixed</b> or retractable writing table that can hold at least 10kgs
11. ADVANCED AND SAFETY FEATURES:  11.3 In case of no power and totally discharged battery, machine must be shifted automatically from	11. ADVANCED AND SAFETY FEATURES:  11.3 In case of no power and totally discharged battery, machine <b>can be shifted from Ventilator</b>

	<p>Republic of the Philippines          Department of Health, Center for Health Development (CHD) IV-CALABARZON  <b>BATANGAS MEDICAL CENTER</b>          Batangas City  <b>ISO 9001:2015 CERTIFIED</b></p>	
---	--	---



Ventilator mode to Manual mode without any switch or button required.	<b>mode to Manual mode automatically or thru a switch or button.</b>
<b>13. PARAMETERS</b>  13.1. Minimum Tidal Volume (Vt) required: 10 to 1600ml  13.2. Breathing Frequency (rate): 1 to 100 bpm (1/min)  13.7. Inspiration/expiration ratio: (Ti:Te) 9:1 to 1:99	<b>13. PARAMETERS</b>  13.1. <b>Tidal Volume (Vt) required: 15ml-1600ml</b>  13. 2. Breathing frequency (rate) : <b>4-60 bpm</b> (1/min)  <b>13.7. To retain</b>
13.10. Built in capnography and SPO2 Monitoring	<b>13.10. Preferably 1 built-in capnography but will consider two (2) Capnography for availability of a backup monitor during lap surgery</b>  <b>Built-in SPO2 is not required to be part of the machine but should be part of the monitor</b>
<b>CARDIAC MONITOR:</b>  8. BUILT IN CAPNOGRAPH (preferably mainstream)	<b>CARDIAC MONITOR:</b>  <b>To retain</b>

**B. IB2021-007B SUPPLY AND DELIVERY OF VARIOUS OFFICE SUPPLIES-REBID**

**ADDITIONAL** Specification as Technical Offer:

**D. Delivery within ten (10) calendar days upon receipt of P.O.**

FROM	TO
5. BAR CODE SCANNER: 4x3 – 20 rolls	5. BAR CODE SCANNER: <ul style="list-style-type: none"> <li>4x3 – 20 rolls</li> <li><b>For Intermec PC43T</b></li> <li><b>Chemical Resistant</b></li> <li><b>Width – 3”</b></li> </ul>
16. ENVELOPE, BROWN A4 SIZE: <ul style="list-style-type: none"> <li>1 BOX</li> </ul> Unit Cost: 675.00	16. ENVELOPE, BROWN A4 SIZE: <ul style="list-style-type: none"> <li><b>500 pc</b></li> </ul> <b>Unit Cost: 1.35</b>
17. ENVELOPE, BROWN LEGAL: <ul style="list-style-type: none"> <li>6 BOXES</li> </ul> Unit Cost: 927.16	17. ENVELOPE, BROWN LEGAL SIZE: <ul style="list-style-type: none"> <li><b>3000 pc</b></li> </ul> <b>Unit Cost: 1.85</b>
18. ENVELOPE, MAILING, WHITE LONG: <ul style="list-style-type: none"> <li>1 BOX</li> </ul>	18. ENVELOPE, MAILING, WHITE LONG: <ul style="list-style-type: none"> <li><b>500 pc</b></li> </ul>

	<p>Republic of the Philippines          Department of Health, Center for Health Development (CHD) IV-CALABARZON  <b>BATANGAS MEDICAL CENTER</b>          Batangas City  <b>ISO 9001:2015 CERTIFIED</b></p>	
---	--	---

Unit Cost: 378.20	Unit Cost: 0.76
39. TAPE, SCOTCH 1 INCH • 100 rolls	<b>39. CLEAR, ADHESIVE TAPE, 1 INCH</b> • 100 rolls
40. TAPE, SCOTCH 2 ICH • 100 pc	<b>40. CLEAR, ADHESIVE TAPE, 2 INCH</b> • 100 pc
41. STAMP PAD • 50 pc	41. STAMP PAD • <b>87x143mm or 3.42 x 5.63 inches</b> • 50 pc
50. TISSUE PAPER 12'S/PACK • 150 pack	50. <b>ROLL</b> TISSUE PAPER 12'S/PACK • 150 pack

- REAM as UNIT of paper quantity must be minimum of 500 sheets

**C. IB2021-007C SUPPLY AND DELIVERY OF VARIOUS HARDWARE MATERIALS FOR REPAIR AND MAINTENANCE OF HOSPITAL BUILDINGS-REBID**

ADDITIONAL Specification as Technical Offer:

**D. Delivery within 15 calendar days upon receipt of initial P.O. Batch Order**

**CARPENTRY SUPPLIES**

	FROM	TO
15	Rivet	Rivet • 1/8x3/4 • 3/16x3/4
23	GI sheet Corrugated	GI sheet Corrugated <b>10 Ft</b>
24	Gutter	Gutter ( <b>Common Ordinary</b> )
25	Screw Assorted	Screw Assorted • <b>Metal Tekscrew 2.5"</b> • <b>Wood Screw 1"</b> • <b>Wood Screw 2"</b>
26	C-Farlings	C-Farlings • <b>2x3</b>
27	Ridge Roll	Ridge Roll • <b>Ordinary (1.2mmx8")</b>
28	Angle Bar, assorted	Angle Bar • <b>1x1x1/4</b> • <b>2x2x1/4</b>
29	Deformed Bar	Deformed Bar





		<ul style="list-style-type: none"> <li>• <b>10mm</b></li> </ul>
30	Flat bar	Flat bar <ul style="list-style-type: none"> <li>• <b>1/4x1"</b></li> </ul>
33	Square Bar	Square Bar <ul style="list-style-type: none"> <li>• <b>12mm</b></li> </ul>
34	G.I. Wire	G.I. Wire <ul style="list-style-type: none"> <li>• <b>G16</b></li> </ul>
35	Common Wire Nail	Common Wire Nail <ul style="list-style-type: none"> <li>• <b>1"</b></li> <li>• <b>2"</b></li> <li>• <b>3"</b></li> <li>• <b>Finishing nail 2"</b></li> </ul>
36	Concrete Nail	Concrete Nail <ul style="list-style-type: none"> <li>• <b>1"</b></li> <li>• <b>2"</b></li> <li>• <b>3"</b></li> </ul>
37	Tubular Bar (assorted)	Tubular Bar (assorted) <ul style="list-style-type: none"> <li>• <b>1x2x1.5</b></li> <li>• <b>2x2x1.5</b></li> </ul>
38	Carrying Channel	Carrying Channel <ul style="list-style-type: none"> <li>• <b>1x2</b></li> </ul>
40	Metal Screw	Metal Screw <ul style="list-style-type: none"> <li>• <b>1"</b></li> <li>• <b>2"</b></li> </ul>
46	GI Pipe	GI Pipe <ul style="list-style-type: none"> <li>• <b>1/2" G 40</b></li> <li>• <b>3/4" G 40</b></li> <li>• <b>1 1/2" G 40</b></li> </ul>
47	BI Pipe	BI Pipe <ul style="list-style-type: none"> <li>• <b>2" X G 40</b></li> </ul>
48	Steel Plate	Steel Plate <ul style="list-style-type: none"> <li>• <b>4 X 8</b></li> </ul>

**ELECTRICAL SUPPLIES**

	FROM	TO
18	Service Entrance Cap	Service Entrance Cap <ul style="list-style-type: none"> <li>• <b>2"</b></li> <li>• <b>3"</b></li> <li>• <b>4"</b></li> </ul>
19	PVC Male Plug	PVC Male Plug <ul style="list-style-type: none"> <li>• <b>1/2</b></li> <li>• <b>3/4</b></li> </ul>



		<ul style="list-style-type: none"> <li>• 1</li> <li>• 1 ¼</li> <li>• 1 ½</li> <li>• 2</li> <li>• 3</li> </ul>
20	PVC Male Plug w/ lock nut	PVC Male Plug w/ lock nut <ul style="list-style-type: none"> <li>• ½</li> <li>• ¾</li> <li>• 1</li> <li>• 1 ¼</li> <li>• 1 ½</li> <li>• 2</li> <li>• 3</li> </ul>
21	PVC Elbow (Assorted)	PVC Elbow (Assorted) <ul style="list-style-type: none"> <li>• ½</li> <li>• ¾</li> <li>• 1</li> <li>• 1 ¼</li> <li>• 1 ½</li> <li>• 2</li> <li>• 3</li> </ul>
22	PVC Clamp (Assorted)	PVC Clamp (Assorted) <ul style="list-style-type: none"> <li>• 1/2</li> <li>• 3/4</li> <li>• 1</li> <li>• 1 ¼</li> <li>• 1 ½</li> <li>• 2 (pack)</li> <li>• 3</li> </ul>
23	RSC Conduit Fittings(Elvee)	RSC Conduit Fittings(Elvee) <ul style="list-style-type: none"> <li>• ½</li> <li>• ¾</li> <li>• 1</li> </ul>
25	Black Screw (Pointed)	Black Screw (Pointed) 1"
26	Pin Light Housing	Pin Light Housing 6"
27	Receptacle Outlet Plastic	Receptacle Outlet Plastic 3"
28	LED Bulb	LED Bulb 9WATTS
29	LED Flood Light	LED Flood Light (50 to 100watts)
30	LED Tube Light	LED Tube Light (T8)
31	LED Tube Light w/ holder	LED Tube Light w/ holder (T8)


	<p>Republic of the Philippines          Department of Health, Center for Health Development (CHD) IV-CALABARZON  <b>BATANGAS MEDICAL CENTER</b>          Batangas City  <b>ISO 9001:2015 CERTIFIED</b></p>	
	 	

34	Exhaust fan	Exhaust fan <ul style="list-style-type: none"> <li>• <b>Duct</b></li> <li>• <b>Ceiling</b></li> </ul>
38	1 Gang Outlet	1 Gang Outlet <b>3 prong (Universal)</b>
39	2 Gang Outlet	2 Gang Outlet <b>3 prong (Universal)</b>
40	Extension outlet	Extension outlet <b>(Universal)</b>
41	Flat Cord	Flat Cord <b>#16</b>
48	Enclosure Circuit Breaker w/ Universal Outlet	Enclosure Circuit Breaker w/ Universal Outlet <ul style="list-style-type: none"> <li>• <b>20 amps</b></li> <li>• <b>30 amps</b></li> <li>• <b>40 amps</b></li> </ul>
51	Circuit Breaker w/ Enclosure 60, 70 amps	Circuit Breaker w/ Enclosure and outlet (with outlet) <ul style="list-style-type: none"> <li>• <b>60 amps</b></li> <li>• <b>70 amps</b></li> </ul>
52	Terminal Lug	Terminal Lug <ul style="list-style-type: none"> <li>• <b>60</b></li> <li>• <b>80</b></li> <li>• <b>125</b></li> <li>• <b>250</b></li> </ul>
54	Secondary Rack	Secondary Rack <b>3 pole</b>

**PLUMBING SUPPLIES**

	FROM	TO
14	Portable Submersible Pump	Portable Submersible Pump <ul style="list-style-type: none"> <li>• <b>2hp with complete fittings</b></li> </ul>
15	Pressure Tank Stainless	Pressure Tank Stainless <ul style="list-style-type: none"> <li>• <b>600 liters</b></li> </ul>
247	Flexible Hose	Flexible Hose <ul style="list-style-type: none"> <li>• <b>1/2 x 12"</b></li> <li>• <b>1/2 x 16"</b></li> </ul>

- **IB2021-007C SUPPLY AND DELIVERY OF VARIOUS HARDWARE MATERIALS FOR REPAIR AND MAINTENANCE OF HOSPITAL BUILDINGS (REBID) is for LINE ITEM BIDDING**



This Supplemental / Bid Bulletin shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 19<sup>th</sup> day of May 2021 in Batangas City.

**(Sgd.) ELIZABETH V. PALINES, MD, FPPS, FCNSP, FPNA**  
Chairperson, HBAC

Received by the bidder:

\_\_\_\_\_  
Signature over printed name

Date Received: \_\_\_\_\_