



HOSPITAL BIDS AND AWARDS COMMITTEE
 SUPPLEMENTAL BID BULLETIN NO. 2
IB 2021-003
 PRE-BID CONFERENCE (APRIL 21, 2021)

- I. Amendment on Terms of Specifications/Terms of Reference as per Pre-Bidding Conference conducted last April 21, 2021 as follows:

IB 2021-003C SUPPLY AND DELIVERY OF 1 UNIT ANESTHESIA MACHINE

FROM	TO
<p>3.1 with at least 8 ventilation modes: VCV, PCV, PRVC, SPONT/PSV and Manual</p> <p>10.2. With a retractable writing table that can hold at least 10kgs.</p> <p>13.10. Built in capnography and SPO2 Monitoring</p> <p>13.2. Breathing Frequency (rate): 1 to 100 bpm (1/min)</p>	<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> • with at least 6 ventilation modes: VCV, PCV, PRVC or SIMV, SPONT/PSV and Manual • With a fixed or retractable writing table that can hold at least 10kgs • preferably built-in capnography but will consider two (2) Capnography for availability of a backup monitor during lap surgery • Built-in SPO2 is not required to be part of the machine but should be part of the monitor • Breathing frequency (rate) : 1-60 bpm (1/min)

SGD. ELIZABETH V. PALINES, MD, FPPS, FCNSP, FPNA
 Chairperson, HBAC

Received by the bidder:

 Signature over printed name

Date Received: _____