


**HOSPITAL BIDS AND AWARDS COMMITTEE**  
**SUPPLEMENTAL BID BULLETIN NO. 1**  
**IB 2022-007 & 008**  
**PRE-BID CONFERENCE (MARCH 29, 2022)**

**I. GENERAL INSTRUCTIONS**

- Bidders are required to submit one (1) soft copy of the original documents clearly scanned in multiple pages in one (1) PDF File per the three required files to the designated Google Form link.
- Documentary requirements on the electronic bidding is the same as in the manual bidding as provided in the 2016 Revised Implementing Rules and Regulations (IRR) of Republic Act 9184 (RA 9184), otherwise known as the “Government Procurement Reform Act” and Resolution No. 09-2020 of the Government Procurement Policy Board entitled “Approving Measures for the Efficient Conduct of Procurement Activities During a State of Calamity, or Implementation of Community Quarantine or Similar Restrictions”
- The BAC is using the 6<sup>th</sup> Edition of the Philippine Bidding Documents and all Amendments Forms therein.
- Due to limited slots in Zoom, only one (1) representative per bidder will be allowed to enter.
- Bidder shall only provide the password of their encrypted file upon being called and acknowledged in the Online Bidding Conference by the BAC before the opening of said Bidder’s bid documents. Bidder’s video should be on during this process.
- For the issuance of the Official Receipt, bidders may get the same at the Cash Collection Section or Procurement Section, during office hours, M-F; 8:00AM-5:00PM, except holidays, and upon showing of proof of payment. For the purpose of the bidding, those who submitted the proof of payment early will be verified by our Cash Collection Section and will be provided a soft copy of the Official Receipt, otherwise, other proof of payment will be acceptable subject to later verification by the Cash Collection Section.
- The winning bidder is required to produce two (2) hard copies of the scanned document anytime before the award. The same can be personally sent or via courier. Any discrepancy on the submitted soft copy as bidder during the bidding conference and hard copies as winning bidder is a ground for disqualification.
- The scanned copy should have the required original signature. E-signature is not allowed.
- Bidders must submit updated Tax clearance per E.O. No. 398, s. 2005, as finally reviewed and approved by the Bureau of Internal Revenue (BIR).
- It is understood that the Contractor is legally responsible to deliver all issued purchase order/s and failure to deliver the first Purchase Order as scheduled shall mean automatic cancellation of the PO and Notice of Award (NOA). Upon cancellation, the BAC shall proceed to qualify the second lowest bidder if applicable; or proceed to Negotiated Procurement. The Winning Contractor who failed to deliver shall shoulder the price difference (from the second lowest bidder) of the item in addition to the acquired liquidated damages.



- In lieu of the Certificate of Ongoing Projects/Accomplishments, Bidders must present their progress report or percentage of partial accomplishment, together with their contracts, awards, and proof of deliveries such as sales invoice.
- Bidders must specify in the Technical Specifications Form under Bidder’s Offer, Technical Specifications Column the exact specifications they are offering submitted against each of the individual parameters. Whereas, bidders must state either “Comply” or “Not Comply” under the Bidder’s Offer, Statement of Compliance Column. The bids with incomplete and incorrect filled out Technical Specifications Form will be disqualified.
- The basis for the computation of the SLCC is the TOTAL ABC of the project. Bidder must have completed, within the period specified in the Invitation to Bid, an SLCC that is similar to the contract to be bid, and whose value, adjusted to current prices using the Philippine Statistics Authority (PSA) consumer price indices, must be at least fifty percent (50%) of the TOTAL ABC OF THE PROJECT. However, in the case of Expendable Supplies, said SLCC must be at least twenty-five percent (25%) of the TOTAL ABC OF THE PROJECT. Service will fall on the Expendable Supplies.
- If there is no single completed contract for the required amount, the bidder may submit at least two (2) completed similar contracts and the aggregate contract amounts should be equivalent to at least the percentage of the ABC as required above and the largest of these similar contracts must be equivalent to at least half of the percentage of the ABC as required.
- A contract shall only be considered “similar” to the contract to be bid if it is of similar nature.
- Contracts similar to the Project should be completed within two (2) years to the deadline for the submission and receipt of bids.
- With regards to the documents supporting the Statement Identifying the Single Largest Completed Contract (SLCC), in lieu of Contract, Certificate of Completion and Certificate of Acceptance, the BAC shall accept Notice of Award (NOA) or Purchase Order (PO), Sales Invoice and Official Receipt (OR), respectively.

II. Amendments/Clarifications on the **Section VII. Technical Specifications** as per Pre-Bidding Conference conducted last March 29, 2022 as follows:

**1. IB2022-008B Supply and Delivery of Various Otolaryngology Equipment**

PARTICULAR	FROM	TO
ITEM NO. 1: VIDEO ENDOSCOPY SYSTEM	<b>TECHNICAL DESCRIPTION:</b>  The set has the following components:  A. Digital Endoscopy Camera HDTV format	<b>TECHNICAL DESCRIPTION:</b>  The set has the following components:  A. Digital Endoscopy Camera HDTV format



Republic of the Philippines  
 Department of Health, Center for Health Development (CHD) IV-CALABARZON  
**BATANGAS MEDICAL CENTER**  
 Batangas City  
**ISO 9001:2015 CERTIFIED**



	<p>3 CCD</p> <p>resolution: 1920 x 1080 pixels or higher</p> <p>with focusing or zoom coupler</p> <p>progressive scan</p> <p>adjustable shutter area</p> <p>freeze and mirror function</p> <p>video output:</p> <p>Y/C video output</p> <p>RGB video output</p> <p>VBS</p> <p>HD-SDI</p> <p>DVI-D</p>	<p>3 CCD or CMOS</p> <p>resolution: 1920 x 1080 pixels or higher</p> <p>with focusing or zoom coupler</p> <p>progressive scan</p> <p>adjustable shutter area</p> <p>freeze and mirror function</p> <p>video output:</p> <p>Y/C video output</p> <p>RGB video output or Y-Pb-Pr</p> <p>VBS</p> <p>HD-SDI</p> <p>DVI-D or HDMI</p>
	<p>B. HD Monitor</p> <p>HD resolution 1920 x 1080 or higher</p> <p>TFT or active matrix display at least 26"</p> <p>display</p> <p>medical grade</p>	<p>B. HD Monitor</p> <p>HD resolution 1920 x 1080 or higher</p> <p>TFT or IPS active matrix display at least 26"</p> <p>display</p> <p>medical grade</p>
	<p>C. HD Video Capture and Storage Device</p> <p>Input/Output: DVI/(RGB with Adapter)</p> <p>HD-SDVDVI (RGB, VGA with Adapter)</p> <p>Image format: JPG, BMP</p> <p>Video format: H.264, MPEG4</p>	<p>C. HD Video Capture and Storage Device</p> <p>Input/Output: <b>HDMI/DVI/SDI/S Video</b> (RGB with Adapter)</p> <p>HD-SDVDVI (RGB, VGA with Adapter)</p> <p>Image format: JPG, BMP, PNG</p> <p>Video format: H.264, MPEG4 (1080,720, PAL, NTSC) or</p>



Republic of the Philippines  
 Department of Health, Center for Health Development (CHD) IV-CALABARZON  
**BATANGAS MEDICAL CENTER**  
 Batangas City  
**ISO 9001:2015 CERTIFIED**





	(1080,720, PAL, NTSC)  USB stick, video cables  1 terabyte external hard drive with cables	<b>MP4(1920x1080 p60)</b>  USB stick, video cables  1 terabyte external hard drive with cables
	E. LED Light Source LED technology  port with standard adaptors for other brands  light intensity adjustment: manual or automatic  automatic thermal protection system  at least 30,000 hours bulb life	D. LED Light Source LED technology  port with standard adaptors for other brands  light intensity adjustment: manual or automatic  automatic thermal protection system  at least <b>20,000</b> hours bulb life
	<b>DOCUMENTATION:</b>  With brochure, manual of operation laser or acetate etching (BatMC ENT-HNS)	<b>DOCUMENTATION:</b>  With brochure, manual of operation laser or acetate etching ( <b>BatMC ORL-HNS</b> )
ITEM NO. 2: <b>OPD LARYNGOLOGY SET</b>	<b>TECHNICAL DESCRIPTION:</b>  3. Rhino-Laryngo-Broncho Fiberscope, flexible (1 set) outer diameter: 3.5-4.0 mm working channel at least 1.3 mm inner diameter 0 degrees direction/angle of view distal tip deflection up/down 100-180 degrees/100-140 degrees a. carrying/storage case b. ETO cap or equivalent c. leakage tester d. flexible double action biopsy forceps at least 60 cm length e. flexible double action universal grasping	<b>TECHNICAL DESCRIPTION:</b>  3. Rhino-Laryngo-Broncho Fiberscope, flexible (1 set) outer diameter: <b>3.4-4.2</b> mm working channel at least <b>1.15</b> mm inner diameter 0 degrees direction/angle of view <b>or 3-50mm field of view</b> distal tip deflection up/down 100-180 degrees/100-140 degrees a. carrying/storage case b. ETO cap or equivalent c. leakage tester d. flexible double action biopsy forceps at least 60 cm length e. flexible double action universal grasping



Republic of the Philippines  
 Department of Health, Center for Health Development (CHD) IV-CALABARZON  
**BATANGAS MEDICAL CENTER**  
 Batangas City  
**ISO 9001:2015 CERTIFIED**



	<p>forceps at least 60 cm length</p> <ul style="list-style-type: none"> <li>f. cap for biopsy channel</li> <li>g. cleaning brush</li> <li>h. mouth piece/guard (4)</li> <li>i. led light source</li> </ul> <ul style="list-style-type: none"> <li>j. UPS at least 1500 watts</li> <li>k. Medical grade monitor at least 24 inches</li> </ul> <ul style="list-style-type: none"> <li>l. with adapters for other brands</li> </ul>	<p>forceps at least 60 cm length</p> <ul style="list-style-type: none"> <li>f. cap for biopsy channel</li> <li>g. cleaning brush</li> <li>h. mouth piece/guard (4)</li> <li>i. led light source</li> </ul> <p><b>Automatic light intensity control</b></p> <p><b>LED daylight quality at least 5,500K lumens</b></p> <ul style="list-style-type: none"> <li>j. UPS at least 1500 watts</li> <li>k. Medical grade monitor at least 24 inches</li> </ul> <p><b>They should be composed of a scratch – and chemical resistant material that can withstand the most common cleaning solutions and disinfectants/with antimicrobial coating or surface treatments that can strengthen infection control.</b></p> <ul style="list-style-type: none"> <li>l. with adapters for other brands</li> </ul>
	<p><b>DOCUMENTATION:</b></p> <p>With brochure, manual of operation laser or acetate etching (BatMC ENT-HNS)</p>	<p><b>DOCUMENTATION:</b></p> <p>With brochure, manual of operation laser or acetate etching (<b>BatMC ORL-HNS</b>)</p>
<p>ITEM NO.  <b>3: ENT OPD        OTOLGY SET</b></p>	<p><b>TECHNICAL DESCRIPTION:</b></p> <p>24. Elevator, Freer (1 pc) stainless steel</p> <p>width: 4.5/4.6 mm</p> <p>over-all length: 180-200 mm</p>	<p><b>TECHNICAL DESCRIPTION:</b></p> <p>24. Elevator, Freer (1 pc) stainless steel</p> <p>width: <b>4.3-4.5</b>/4.6 mm</p> <p>over-all length: 180-200 mm</p>

	<p>Republic of the Philippines          Department of Health, Center for Health Development (CHD) IV-CALABARZON  <b>BATANGAS MEDICAL CENTER</b>          Batangas City  <b>ISO 9001:2015 CERTIFIED</b></p>	
---	--	---

**2. IB2022-008C Supply and Delivery of Various Ophthalmology Equipment**

PARTICULAR	FROM	TO
ITEM NO. <b>2: PHOTOCOAGULATING/ 532 LASER</b>	<b>TECHNICAL DESCRIPTION:</b>  POWER OUTPUT ON CORNEA: at least 2W	<b>TECHNICAL DESCRIPTION:</b>  POWER OUTPUT ON CORNEA: <b>at least 1.5W</b>
	With pattern laser delivery system  With optional voice confirmation & remote control  With wireless foot switch control  With integrated CSO Slit lamp, Zeiss-type  With cooling system	With pattern laser delivery system  With optional voice confirmation & remote control  With wireless foot switch control  With integrated CSO Slit lamp, Zeiss-type, <b>or Haag streit Type or their equivalent</b>  With cooling system
	<b>PHYSICAL DESCRIPTION:</b>  Dimensions: at least 12 in (30 cm) x 8.5 in (21 cm) x 14 in (35 cm)  Weight (lbs, kg): at least 3kg  Noise (in dBA): NA  Heat dissipation: NA  Mobility, portability: NA	<b>PHYSICAL DESCRIPTION:</b>  Dimensions: at least 12 in (30 cm) x <b>5 in (21 cm)</b> x 14 in (35 cm)  Weight (lbs, kg): at least 3kg  Noise (in dBA): NA  Heat dissipation: NA  Mobility, portability: NA
	<b>ACCESSORIES, CONSUMABLES, SPARE PARTS. OTHER COMPONENT:</b>  1. Laser Contact Lens PRP 160  2. WideField lens  3. Mainster or Superquad Lens  4. Micro pulse laser trabeculoplasty lens	<b>ACCESSORIES, CONSUMABLES, SPARE PARTS. OTHER COMPONENT:</b>  1. Laser Contact Lens PRP 160 or <b>Laser Contact Lens PRP 165</b> <b>2. Mainster WideField lens or its equivalent</b> <b>3. Mainster or Superquad Lens 160 or their equivalent</b> <b>4. Micro pulse laser trabeculoplasty lens or its equivalent</b>



Republic of the Philippines  
 Department of Health, Center for Health Development (CHD) IV-CALABARZON  
**BATANGAS MEDICAL CENTER**  
 Batangas City  
**ISO 9001:2015 CERTIFIED**



	5. Motorized Table 6. UPS System 7. Stools(2 pcs)	5. Motorized Table 6. UPS System 7. Stools(2 pcs)
	<p><b>ENVIRONMENTAL REQUIREMENTS:</b></p> <p>Operating Condition: Capable of operating continuously in ambient temperature of 5 to 40 deg C and relative humidity of 15 to 80% in ideal circumstances</p> <p>Storage condition: Capable of being stored continuously in ambient temperature of 0 to 40 deg C and relative humidity of 15 to 90%</p> <p>Sterilization not required</p>	<p><b>ENVIRONMENTAL REQUIREMENTS:</b></p> <p>Operating Condition: Capable of operating continuously in ambient temperature of 5 to 40 deg C and relative humidity of 15 to <b>90%</b> in ideal circumstances</p> <p>Storage condition: Capable of being stored continuously in ambient temperature of <b>0 to 60 deg C</b> and relative humidity of 15 to 90%</p> <p>Sterilization not required.</p>

This Supplemental / Bid Bulletin shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

**(Sgd.) MERLITA G. PUBLICO, MD, FPAFP, MHA**  
 Chairperson, HBAC

Received by the bidder:

\_\_\_\_\_  
 Signature over printed name  
 Date Received: \_\_\_\_\_