



Republic of the Philippines
Department of Health, Center for Health Development (CHD) IV-CALABARZON
BATANGAS MEDICAL CENTER
Batangas City
ISO 9001:2015 CERTIFIED



HOSPITAL BIDS AND AWARDS COMMITTEE
SUPPLEMENTAL BID BULLETIN NO. 2
IB2022-008B Supply & Delivery of Various Otolaryngology Equipment

This Supplemental/Bid Bulletin No. 2 is being issued to inform all of the **REVISED Technical Specifications** specifically of **Item No. 2 – OPD Laryngology Set** as follows:

ITEM NO. 2 – OPD LARYNGOLOGY SET

ABC: PHP 3,000,000.00

QUANTITY: 1 COMPLETE SET

PURPOSE OF USE:

This set is essential in the examination of the pharyngeal and laryngeal areas as well as procedures that can be done in an out-patient basis

TECHNICAL DESCRIPTION:

1. *Biopsy Forceps (1pc)
double action jaws, flexible
1mm diameter, 60-150 cm length*
2. *Laryngeal Mirrors (10pcs)
stainless steel, autoclavable
fixed handle
size 6, 22 mm*
3. *Rhino-Laryngo-Broncho Fiberscope, flexible (1set) outer diameter: 3.4-4.2mm
working channel at least 1.15mm inner diameter
0 degrees direction/angle of view or 3-50mm
field of view
distal tip deflection up/down
100-180 degrees/100-140 degrees*
 - a. *carrying/storage case*
 - b. *ETO cap or equivalent*
 - c. *leakage tester*
 - d. *flexible double action biopsy forceps
at least 60 cm length*
 - e. *flexible double action universal grasping
forceps at least 60 cm length*
 - f. *cap for biopsy channel*
 - g. *cleaning brush*
 - h. *mouth piece/guard (4)*
 - i. **led light source**
 - **stand alone or integrated**



Republic of the Philippines
Department of Health, Center for Health Development (CHD) IV-CALABARZON
BATANGAS MEDICAL CENTER
Batangas City
ISO 9001:2015 CERTIFIED



j. camera system with recorder

k. UPS at least 1500 watts

l. Medical grade monitor at least **18** inches

- They should be composed of a scratch – and chemical resistant material that can withstand the most common cleaning solutions and disinfectants/with antimicrobial coating or surface treatments that can strengthen infection control.

m. with adapters for other brands

4. Fiber Optic Light Cable (1pc)

diameter: at least 3.5 mm

length: at least 2.3 m

with adapters for other brands

PHYSICAL DESCRIPTION:

Refer to technical description

UTILITY REQUIREMENTS:

Voltage: 220-240VAC 60Hz or auto volt for light source

LED at least 5500K lumens light source

Uninterrupted power supply of at least 1500 watts

ACCESSORIES, CONSUMABLES, SPARE PARTS, OTHER COMPONENT:

a. cleaning brush (5 pcs)

b. mouth piece/guard (5 pcs)

ENVIRONMENTAL REQUIREMENTS:

none

DELIVERY:

Must be delivered within 60-90 calendar days from receipt of Purchase Order. In case the unit is not yet available, supplier must furnish a service unit. Failure to provide shall automatically terminate the contract, in addition to other penalties applicable. The supplier shall provide service unit within the warranty period in case of breakdown. Before delivery, supplier must call the Materials Management Section for delivery coordination to avoid non-acceptance or delay in acceptance.

TRAINING, INSTALLATION & UTILIZATION:

Training for proper care and storage

Maintenance every quarter for first 2 years, yearly thereafter

Manufacturer/supplier to perform installation, safety and operation checks, and preventive maintenance



Republic of the Philippines
Department of Health, Center for Health Development (CHD) IV-CALABARZON
BATANGAS MEDICAL CENTER
Batangas City
ISO 9001:2015 CERTIFIED



before handover. Acceptance tests to be specified and local clinical and technical staff to verify proper and full functioning of device.

Training of users in operation shall be provided. Training of medical equipment technician personnel on basic maintenance and troubleshooting also to be provided by supplier or manufacturer.

Information to be provided by manufacturer/supplier, e.g. cleaning, disinfection/ sterilization method (for reusable devices).

WARRANTY:

At least 2 years warranty period

AT LEAST 2 YEARS WARRANTY PERIOD ON PARTS AND SERVICES WITH CERTIFICATE (INCLUSION AND EXCLUSION TO BE SPECIFIED)

SUPPLIER TO PERFORM (QUARTERLY) PREVENTIVE MAINTENANCE SERVICE WITHIN THE WARRANTY PERIOD WITH PM STICKER AND SERVICE REPORT

PREVENTIVE MAINTENANCE SCHEDULE/CALENDAR TO BE PROVIDED BY SUPPLIER

DOCUMENTATION:

*With brochure, manual of operation
laser or acetate etching (BatMC ORL-HNS)*

SAFETY & STANDARD:

USFDA approved or CE / EN marked

ISO 13485

This Supplemental / Bid Bulletin shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

(Sgd.) MERLITA G. PUBLICO, MD, FPAFP, MHA
Chairperson, Bids and Awards Committee