

SUPPLEMENTAL BID BULLETIN
Addendum No. 2016-007

This Addendum No. 2016-007 is issued to modify or amend items in the Bid Documents. This shall form an integral part of the Bid Documents.

1. PROJECT: ITB 2016-009A

Procurement of Outsourced Pulmonary Unit Mechanical Ventilator (Rebid)


- **Specifications (both for Volume Cycled (Adult/Pediatric) Ventilators and Pressure Control (Infant/Neonate) Ventilators)**
 - Machine not older than 10 years from manufacturing date
- **Maintenance and Servicing**
 - Change of filter must be done once a week or after every change of patient

2. PROJECT: ITB 2016-009A / ITB 2016- 009B

Procurement of Outsourced Pulmonary Unit Mechanical Ventilator (Rebid)

Procurement of Drugs And Medicines (rebid)

- ***NFCC computation***
 - the NFCC's K factor shall be fixed at **15**, regardless of contract duration


ELIZABETH V. PALINES, MD, FPPS, FCNSP, FPNA
BAC Chairperson

Received by the bidder:

Signature over printed name

Date Received: _____