



Terms of Reference for the Outsourcing of Endoscopy Machine

1. SPECIFICATIONS:

- With HD Video Processor and CLV
- Light Source with Pigtail, Keyboard, Light Control Cable
- Water Bottle, RGB Cable and Power Cords
 - ✓ With Video Colonoscope with 12.8mm Outer Diameter
 - ✓ 3.7mm Working Channel
 - ✓ 170 degree Angle of View, 168cm Working Length
 - ✓ Angulation 180°/180° Up/Down and 160°/160° Right/Left
 - ✓ With Video Gastroscope with 8.8mm Outer Diameter
 - ✓ 2.8mm Working Channel
 - ✓ 140° Angle of View, 103cm Working Length, Angulation 210°/90° Up/Down and 100°/100° Right Left
- With Medical Grade Monitor with Stand
- With Procedure Cart
- Biopsy Forcep 230cm in length
- Cleaning Brush

2. MAINTENANCE AND SERVICE

- ✓ The winning provider shall ensure the availability of qualified technician and trained personnel to maintain the equipment.
- ✓ Shall submit the Schedule of corrective maintenance upon facilitation of agreement.
- ✓ The supplier shall provide one unit and shall provide additional unit for back-up if the if needed.

3. OTHER PROVISION

The proposal shall be evaluated by the end-user for consideration.

Noted by:

FOR 
Dr. Donnazon Reyes-Macasaet