

## Republic of the Philippines Department of Health, Center for Health Development (CHD) IV-CALABARZON

## BATANGAS MEDICAL CENTER

Batangas City

## ISO 9001:2015 CERTIFIED





## HOSPITAL BIDS AND AWARDS COMMITTEE SUPPLEMENTAL BID BULLETIN IB2019-005 NO.2

- I. Amendment of Terms of Specifications as follows:
  - A. IB 2019-005A PROCUREMENT OF MEDICAL AND SURGICAL SUPPLIES

1. ITEM No. 91 - IV CANNULA G16

FROM	ТО
• IV CANNULA G16,	• IV CANNULA G16, ,L:1.7MM/45MM
,L:1.7MM/45MM, STERILE,	to 1.8MM/45MM, STERILE, MADE
MADE OF BIO-COMPATIBLE,	OF BIO-COMPATIBLE,
POLYURETHANE MATERIAL,	POLYURETHANE MATERIAL, NON
NON-PYROGENIC, WITH	PYROGENIC, WITH RADIOPAQUE
RADIOPAQUE INJECTION	INJECTION PORT FLOW: 200ML/
PORT FLOW: 200ML/ MIN	MIN

2. ITEM No. 110 - POLYGLACTIC/POLYGLYCOLIDECO - LACTIDE

FROM	ТО
• POLYGLACTIN/ POLYGLYCOLIDECO- L-LACTIDE, 5-0 ATRAUMATIC ROUND, LENGTH: 70CM (28"), NEEDLE: 13MM 1/2 CIRCLE, MULTIFILAMENT, STERILE, ABSORBABLE, BRAIDED, SMOOTH KNOT RUN DOW	• POLYGLACTIN/ POLYGLYCOLIDECO- L-LACTIDE, 5-0 ATRAUMATIC ROUND, LENGTH: AT LEAST 70CM (28"), NEEDLE: 13MM to 35MM 1/2 CIRCLE, MULTIFILAMENT, STERILE, ABSORBABLE, BRAIDED, SMOOTH KNOT RUN DOW

ELIZABETH V. PALINES, MD, FPPS, FCNSP, FPNA Chairperson, HBAC

Received by the bidder:

Signature over printed name

Date Received: