



Republic of the Philippines
Department of Health, Center for Health Development (CHD) IV-CALABARZON
BATANGAS MEDICAL CENTER
Batangas City
ISO 9001:2015 CERTIFIED



HOSPITAL BIDS AND AWARDS COMMITTEE
SUPPLEMENTAL BID BULLETIN NO. 1
IB 2020-005
PRE-BID CONFERENCE (AUGUST 20, 2020)

I. GENERAL INSTRUCTIONS

- The Batangas Medical Center will only accept bid documents submitted electronically to the designated Google Form link to be given upon showing of proof of payment before the opening date of the bids. Bidders are required to show proof of payment submitted to bac@batmc.doh.gov.ph before it will be given the link for submission of bid documents and link to the Zoom meeting.
- Payment can be made by paying directly to the cashier of the Procuring Entity or through bank transfer at: Bank Name: Land Bank of the Philippines, Account Name: Batangas Medical Center, Account Number: 0302 1161 99. Either way, proof of receipt must be emailed to the aforementioned address to be given the link for submission of bids and link to the Zoom meeting.
 1. Bidders must submit 3 files
 - a) Copy of Official Receipt or Other Proof of Payment
 - b) Eligibility/Technical File
 - c) Financial File
 2. Documents must be continuous scan or multiple pages in one PDF file per the three required files. Include table of contents in the front page to facilitate for quick search of the document.
 3. Files b and c should be individually compressed and password protected using third party applications such as win.zip, win.rar, etc... (Password shall not be disclosed until the day of bid opening)
 4. Use the following format for filename for easy identification of your files.
 - a) ProjectCode_CompanyName_OR
 - b) ProjectCode_CompanyName_Eligibilitydocs
 - c) ProjectCode_CompanyName_Financialdocs
 5. Link to the submission of documents shall be given, emailed or sent only to bidders upon submission of proof of payment.
 6. Documents must be submitted any time before 9:00AM of date of opening of bid and upon showing of proof of payment (Note: Needs time for secretariat to download and organize files before Bid Opening)
 7. Late submission shall be invalidated based on the automatic date/time of submission as recorded in the online facility.
 8. Corrupted or files that are inoperable during the opening are likewise invalidated. Please check your files accordingly after submission.
 9. In case of submission of more than 1 set of files, the latest submission within the set deadline shall be deemed valid.
- Bidders are required to submit one (1) soft copy of the original documents scanned in multiple pages in one (1) PDF File per the three required files to the designated Google Form link.
- Documentary requirements on the electronic bidding is the same as in the manual bidding as provided in the 2016 Revised Implementing Rules and Regulations (IRR) of Republic Act 9184



(RA 9184), otherwise known as the “Government Procurement Reform Act” and Resolution No. 09-2020 of the Government Procurement Policy Board entitled “Approving Measures for the Efficient Conduct of Procurement Activities During a State of Calamity, or Implementation of Community Quarantine or Similar Restrictions”

- Due to limited slots in Zoom, only one (1) representative per bidder will be allowed to enter.
- Bidder shall only provide the password of their encrypted file upon being called and acknowledged in the Online Bidding Conference by the BAC before the opening of said Bidder’s bid documents. Bidder’s video should be on during this process.
- For the issuance of the Official Receipt, bidders may get the same at the Cash Collection Section or Procurement Section, during office hours, M-F; 8:00AM-5:00PM, except holidays, and upon showing of proof of payment. For the purpose of the bidding, those who submitted the proof of payment early will be verified by our Cash Collection Section and will be provided a soft copy of the Official Receipt, otherwise, other proof of payment will be acceptable subject to later verification by the Cash Collection Section.
- In case Certifications requiring authentication by Phil Consulate from Country of origin is not yet available in time of Opening of Bids, Bidder must submit it during Post Qualification.
- The scanned copy should have the required original signature. E-signature is not allowed. The winning bidder is required to produce two (2) hard copies of the scanned document anytime before the award. The same can be personally sent or via courier. Any discrepancy on the submitted soft copy as bidder during the bidding conference and hard copies as winning bidder is a ground for disqualification.
- Document with Apostille Certification is acceptable provided that the country where the document was issued is a party to the Apostille Convention.
- Bidder must have completed, within the period specified in the Invitation to Bid, an SLCC that is similar to the contract to be bid, and whose value, adjusted to current prices using the Philippine Statistics Authority (PSA) consumer price indices, must be at least fifty percent (50%) of the ABC. However, in the case of Expendable Supplies, said SLCC must be at least twenty five percent (25%) of the ABC. The prospective bidder should have completed at least two (2) similar contracts and the aggregate contract amounts should be equivalent to at least the required percentage of the ABC. Service will fall on the Expandable Supplies.
- A contract shall only be considered “similar” to the contract to be bid if it is of similar nature.
- Bidders must not only state in the Technical Specifications Form either “Comply” or “Not Comply” but are also required to specify the exact specifications they are offering and must indicate the reference page number in the user manual submitted against each of the individual parameters. The bids with incompletely filled out tech specs form will be disqualified.

II. Amendment of Terms of Specifications / Terms of Reference as per the Pre-Bidding Conference conducted last August 22, 2020 are as follows:

A. IB 2020-005A SUPPLY AND DELIVERY OF 1 UNIT MRI COMPATIBLE ANESTHESIA MACHINE

FROM	TO
<p><u>I. Anesthesia Machine</u> Features:</p> <ul style="list-style-type: none"> • Electronic touch screen ventilator 	<ul style="list-style-type: none"> • Pneumatic type, touch screen ventilator

	Republic of the Philippines Department of Health, Center for Health Development (CHD) IV-CALABARZON BATANGAS MEDICAL CENTER Batangas City ISO 9001:2015 CERTIFIED	
--	---	--

○ SPECIFICATIONS: <ul style="list-style-type: none"> Approved for use with a 3.0 Tesla scanner, up to the 1000 gauss line 	<ul style="list-style-type: none"> Approved for use with a 3.0 Tesla scanner, minimum of 1000 gauss line
○ SPECIFICATIONS: <ul style="list-style-type: none"> Width: 700-730mm 	<ul style="list-style-type: none"> Width: 700-800mm
2C. Service Unit:	2C. Service Unit: In case of Unit Malfunction, Bidder must provide a service unit within 48 hours until such time that the unit has been repaired or replaced.
	Bidder must have 24/7 readily available Technician to provide technical service and support within 24 hours from report via phone call or email in case of machine breakdown or as needs arise
3B. Preventive Maintenance	Bidder must conduct Quarterly Preventive Maintenance within the warranty period
	Bidder must provide training to the end-users once a year within the warranty period
7. Others a. Certification from the manufacturer authenticated by the Philippine Consulate from the country of origin of the unit that the warranty should not be affected with the change of distributor	7. Others a. Certification from manufacturer authenticated by Phil Consulate from Country of origin of the unit that the warranty should not be affected with change of distributor. In case said Certification is not available in time of Opening of Bids, Bidder must submit it during Post Qualification.
OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB: <ul style="list-style-type: none"> USB 1 should contain the Technical Components of the offering USB 2 should contain the Financial Component of the offering 	For omission

B. IB 2020-005B SUPPLY AND DELIVERY OF OR RHINOLOGY SET



FROM	TO
<ul style="list-style-type: none"> 1 pc - Scope case - plastic, capacity 2 rigid scopes diameter 1-4 mm, over-all length:130 mm 	<ul style="list-style-type: none"> Scope case: can fit in any configuration: <ol style="list-style-type: none"> 1 scope per case 2 scopes per case 3 scopes per case 4 scopes per case 5 scopes per case 6 scopes per case; as long as each scope is accommodated in a scope case
	<ul style="list-style-type: none"> Laser / Acetate Etching USFDA approved, CE/EN marked
1. Power Requirement for Electrical/Electronic unit:	For omission



Republic of the Philippines
 Department of Health, Center for Health Development (CHD) IV-CALABARZON
BATANGAS MEDICAL CENTER
 Batangas City
ISO 9001:2015 CERTIFIED



<p>A. Voltage: 220 – 240 volts B. Frequency: 60 Hz C. Power Cord: Dual Insulated cable, at least 2 meters D. Plug: 3 prong male/female plug, Hospital Grade E. Grounding cable for the equipment with secondary grounding system. “(NOTE: PROVIDE AND ATTACH IMAGES TO PROVE THAT THE ABOVE MENTIONED ARE COMPLIED)”</p>	
<p>3. Acceptance and Maintenance: A. During acceptance: The supplier should conduct actual testing (Qualitative and Quantitative Test) using calibrated analyzers or testing devices. Witness by the end users and the technical inspectors.</p> <ol style="list-style-type: none"> a. Valid certificates of the Technicians/Engineers to conduct service/maintenance b. Valid certificates of calibration of the analyzer and testing equipment c. Provide Service Report per unit d. Calibration Certificates or equivalent e. Acceptance and preventive maintenance Sticker f. Preventive Maintenance Calendar <p>“(NOTE: PROVIDE AND ATTACH DOCUMENT TO PROVE THAT THE ABOVE MENTIONED ARE COMPLIED SUCH as the ff.)”</p> <ol style="list-style-type: none"> a. List of the Engineers/Technicians with their certificates to conduct service and maintenance b. List of the Analyzers/Testing tools with their Brand/Model/Serial No. and its valid certificate of calibration” c. Sample/Template of Service report and Calibration certificate or equivalent d. Sample/Template of Preventive Maintenance Sticker e. Sample/Template of Preventive Maintenance Calendar 	<p>3. Acceptance and Maintenance: A. During acceptance: The supplier should provide the end users and the technical inspectors:</p> <ol style="list-style-type: none"> a. Valid certificates of the Technicians/Engineers to conduct service/ maintenance b. Provide Service Report per unit c. Acceptance and preventive maintenance Sticker d. Preventive Maintenance Calendar <p>“(NOTE: PROVIDE AND ATTACH DOCUMENT TO PROVE THAT THE ABOVE MENTIONED ARE COMPLIED SUCH as the ff.)”</p> <ol style="list-style-type: none"> a. List of the Engineers/Technicians with their certificates to conduct service and maintenance b. Sample/Template of Preventive Maintenance Sticker c. Sample/Template of Preventive Maintenance Calendar
<p>B. Preventive Maintenance</p> <ol style="list-style-type: none"> a. Conduct cleaning and testing all parameters including all accessories. b. Conduct Qualitative and Quantitative Test c. Conduct Calibration if needed, d. Provide ServiceReport per unit. e. Calibration Certificates or equivalent f. Report Findings, Suggestion and Recommendation 	<p>B. Preventive Maintenance</p> <ol style="list-style-type: none"> a. Conduct cleaning and testing all parameters including all accessories. b. Provide Service Report per unit. c. Report Findings, Suggestion and Recommendation
<p>6. Certification that the bidder should provide 2 copies of the following manuals (English Version):</p>	<p>6. Certification that the bidder should provide 2 copies of the Operators Manuals, English Version</p>

	Republic of the Philippines Department of Health, Center for Health Development (CHD) IV-CALABARZON BATANGAS MEDICAL CENTER Batangas City ISO 9001:2015 CERTIFIED	
--	---	--



a. Operators Manuals (attached image of front and table of contents) b. Service Manuals (attached image of front and table of contents) c. Quick Guide attached with the unit (attached image)	(attached image of front and table of contents). If Operators Manual is not available, Instruction for Proper Care and Storage of Instruments may suffice.
	Bidder must have 24/7 readily available Technician to provide technical service and support within 24 hours from report via phone call or email
3B. Preventive Maintenance	Bidder must conduct Quarterly Preventive Maintenance within the warranty period
d. For demonstration/confirmation of specification during the evaluation.	d. For demonstration/confirmation of specification during the evaluation. For manufacturers that has previously supplied instruments to the BatMC or those who have supplied instruments to other DOH-retained hospitals, a product catalogue and few sample instruments shall be required instead of a demonstration.
OTHER TERMS AND CONDITIONS: 1. Copy of End User Acceptance Certificate from 1 government and 1 private institution 2. Training should be once a year for 2 years 3. Preventive and corrective maintenance scheduled quarterly.	OTHER TERMS AND CONDITIONS: 1. At least two (2) End User Acceptance Certificate preferably from 1 government and 1 private institution. 2. Training for proper use, maintenance and storage of instruments for Doctors and Nursing Staff should be once a year for 2 years 3. Preventive and corrective maintenance scheduled quarterly.
OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB: <ul style="list-style-type: none"> • USB 1 should contain the Technical Components of the offering • USB 2 should contain the Financial Component of the offering 	For omission

C. IB 2020-005C SUPPLY AND DELIVERY OF OR LARYNGOSCOPE SET

FROM	TO
<ul style="list-style-type: none"> • 1 pc - Chest Support with Rod - for children, length: 240 mm, rubber ring diameter: at least 90 mm • 1 pc - Chest Support with Rod - overall length: 120 mm, diameter: 1.0 mm, stainless steel 	<ul style="list-style-type: none"> • 1 pc - Chest Support with Rod for Children • 1 pc - Chest Support with Rod for Adults
<ul style="list-style-type: none"> • 1 pc - Scope Case - for tele-laryngoscope 	<ul style="list-style-type: none"> • Scope case for each Tele-Laryngoscope
	<ul style="list-style-type: none"> • Laser / Acetate Etching • USFDA approved, CE/EN marked






<p>2. Power Requirement for Electrical/Electronic unit:</p> <ul style="list-style-type: none"> A. Voltage: 220 – 240 volts B. Frequency: 60 Hz C. Power Cord: Dual Insulated cable, at least 2 meters D. Plug: 3 prong male/female plug, Hospital Grade E. Grounding cable for the equipment with secondary grounding system. <p>“(NOTE: PROVIDE AND ATTACH IMAGES TO PROVE THAT THE ABOVE MENTIONED ARE COMPLIED)”</p>	<p>For omission</p>
<p>2C. Service Unit:</p>	<p>2C. Service Unit: In case of Unit Malfunction, Bidder must provide a service unit within 48 hours until such time that the unit has been repaired or replaced.</p>
<p>3. Acceptance and Maintenance:</p> <ul style="list-style-type: none"> A. During acceptance: The supplier should conduct actual testing (Qualitative and Quantitative Test) using calibrated analyzers or testing devices. Witness by the end users and the technical inspectors. <ul style="list-style-type: none"> a. Valid certificates of the Technicians/Engineers to conduct service/maintenance b. Valid certificates of calibration of the analyzer and testing equipment c. Provide Service Report per unit d. Calibration Certificates or equivalent e. Acceptance and preventive maintenance Sticker f. Preventive Maintenance Calendar <p>“(NOTE: PROVIDE AND ATTACH DOCUMENT TO PROVE THAT THE ABOVE MENTIONED ARE COMPLIED SUCH as the ff.)”</p> <ul style="list-style-type: none"> a. List of the Engineers/Technicians with their certificates to conduct service and maintenance b. List of the Analyzers/Testing tools with their Brand/Model/Serial No. and its valid certificate of calibration)” c. Sample/Template of Service report and Calibration certificate or equivalent d. Sample/Template of Preventive Maintenance Sticker e. Sample/Template of Preventive Maintenance Calendar 	<p>3. Acceptance and Maintenance:</p> <ul style="list-style-type: none"> A. During acceptance: The supplier should provide the end users and the technical inspectors: <ul style="list-style-type: none"> a. Valid certificates of the Technicians/Engineers to conduct service/ maintenance b. Provide Service Report per unit c. Acceptance and preventive maintenance Sticker d. Preventive Maintenance Calendar <p>“(NOTE: PROVIDE AND ATTACH DOCUMENT TO PROVE THAT THE ABOVE MENTIONED ARE COMPLIED SUCH as the ff.)”</p> <ul style="list-style-type: none"> a. List of the Engineers/Technicians with their certificates to conduct service and maintenance b. Sample/Template of Preventive Maintenance Sticker c. Sample/Template of Preventive Maintenance Calendar
<p>B. Preventive Maintenance</p> <ul style="list-style-type: none"> a. Conduct cleaning and testing all parameters including all accessories. b. Conduct Qualitative and Quantitative Test c. Conduct Calibration if needed, 	<p>B. Preventive Maintenance</p> <ul style="list-style-type: none"> a. Conduct cleaning and testing all parameters including all accessories. b. Provide Service Report per unit. c. Report Findings, Suggestion and Recommendation

	<p>Republic of the Philippines Department of Health, Center for Health Development (CHD) IV-CALABARZON BATANGAS MEDICAL CENTER Batangas City ISO 9001:2015 CERTIFIED</p>	
--	--	--

<p>d. Provide ServiceReport per unit. e. Calibration Certificates or equivalent f. Report Findings, Suggestion and Recommendation</p>	
<p>6. Certification that the bidder should provide 2 copies of the following manuals (English Version):</p> <p>a. Operators Manuals (attached image of front and table of contents) b. Service Manuals (attached image of front and table of contents) c. Quick Guide attached with the unit (attached image)</p>	<p>6. Certification that the bidder should provide 2 copies of the Operators Manuals, English Version (attached image of front and table of contents). If Operators Manual is not available, Instruction for Proper Care and Storage of Instruments may suffice.</p>
	<p>Bidder must have 24/7 readily available Technician to provide technical service and support within 24 hours from report via phone call or email</p>
<p>3B. Preventive Maintenance</p>	<p>Bidder must conduct Quarterly Preventive Maintenance within the warranty period</p>
<p>d. For demonstration/confirmation of specification during the evaluation.</p>	<p>d. For demonstration/confirmation of specification during the evaluation.For manufacturers that has previously supplied instruments to the BatMC or those who have supplied instruments to other DOH-retained hospitals, a product catalogue and few sample instruments shall be required instead of a demonstration.</p>
<p>OTHER TERMS AND CONDITIONS:</p> <ol style="list-style-type: none"> 1. Copy of End User Acceptance Certificate from 1 government and 1 private institution 2. Training should be once a year for 2 years 3. Preventive and corrective maintenance scheduled quarterly. 	<p>OTHER TERMS AND CONDITIONS:</p> <ol style="list-style-type: none"> 1. At least two (2) End User Acceptance Certificate preferably from 1 government and 1 private institution. 2. Training for proper use, maintenance and storage of instruments for Doctors and Nursing Staff should be once a year for 2 years 3. Preventive and corrective maintenance scheduled quarterly.
<p>OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB:</p> <ul style="list-style-type: none"> • USB 1 should contain the Technical Components of the offering • USB 2 should contain the Financial Component of the offering 	<p>For omission</p>

D. IB 2020-005D SUPPLY AND DELIVERY OF BRONCHOSCOPE SET (ADULT AND PEDIA)

FROM	TO
	<ul style="list-style-type: none"> • Laser / Acetate Etching USFDA approved, CE/EN marked
<ol style="list-style-type: none"> 1. Copy of End User Acceptance Certificate from 1 government and 1 private 	<ol style="list-style-type: none"> 1. At least 2 End User Acceptance Certificate preferably from 1 government and 1 private

	Republic of the Philippines Department of Health, Center for Health Development (CHD) IV-CALABARZON BATANGAS MEDICAL CENTER Batangas City ISO 9001:2015 CERTIFIED	 
--	---	--

institution d. For demonstration/confirmation of specification during the evaluation.	institution. d. For demonstration/confirmation of specification during the evaluation. For manufacturers that have previously supplied instruments to the BatMC or those who have supplied instruments to other DOH-retained hospitals, a product catalogue and few sample instruments shall be required instead of a demonstration.
OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB: <ul style="list-style-type: none"> • USB 1 should contain the Technical Components of the offering • USB 2 should contain the Financial Component of the offering 	For omission

E. IB2020-005E SUPPLY AND DELIVERY OF 2 UNITS ENT TREATMENT UNIT

FROM	TO
	<ul style="list-style-type: none"> • Laser / Acetate Etching USFDA approved, CE/EN marked
Suction system <ul style="list-style-type: none"> • 110-240v 	Suction system <ul style="list-style-type: none"> • Standard (220-240V) or Autovolt (110-240V)
ENT Examination chair <ul style="list-style-type: none"> • chair rotation at least 360 degrees 	<ul style="list-style-type: none"> • electronic • chair rotation at least 340 degrees
Standard Requirements: 1. Power Requirement for Electrical/Electronic unit: E. Grounding cable for the equipment with secondary grounding system.	For omission
3. Acceptance and Maintenance: A. During acceptance: The supplier should conduct actual testing (Qualitative and Quantitative Test) using calibrated analyzers or testing devices. Witness by the end users and the technical inspectors. <ul style="list-style-type: none"> a. Valid certificates of the Technicians/Engineers to conduct service/maintenance b. Valid certificates of calibration of the analyzer and testing equipment c. Provide Service Report per unit d. Calibration Certificates or equivalent e. Acceptance and preventive maintenance Sticker f. Preventive Maintenance Calendar <p>“(NOTE: PROVIDE AND ATTACH DOCUMENT TO PROVE THAT THE ABOVE MENTIONED ARE COMPLIED SUCH as the ff.)”</p>	3. Acceptance and Maintenance: A. During acceptance: The supplier should provide the end users and the technical inspectors: <ul style="list-style-type: none"> a. Valid certificates of the Technicians/Engineers to conduct service/maintenance b. Provide Service Report per unit c. Acceptance and preventive maintenance Sticker d. Preventive Maintenance Calendar <p>“(NOTE: PROVIDE AND ATTACH DOCUMENT TO PROVE THAT THE ABOVE MENTIONED ARE COMPLIED SUCH as the ff.)”</p> <ul style="list-style-type: none"> a. List of the Engineers/Technicians with their certificates to conduct service and maintenance b. Sample/Template of Preventive

<p>a. List of the Engineers/Technicians with their certificates to conduct service and maintenance</p> <p>b. List of the Analyzers/Testing tools with their Brand/Model/Serial No. and its valid certificate of calibration”</p> <p>c. Sample/Template of Service report and Calibration certificate or equivalent</p> <p>d. Sample/Template of Preventive Maintenance Sticker</p> <p>e. Sample/Template of Preventive Maintenance Calendar</p>	<p>Maintenance Sticker</p> <p>c. Sample/Template of Preventive Maintenance Calendar</p>
<p>d. For demonstration/confirmation of specification during the evaluation.</p>	<p>d. For demonstration/confirmation of specification during the evaluation.For manufacturers that have previously supplied instruments to the BatMC or those who have supplied instruments to other DOH-retained hospitals, a product catalogue and few sample instruments shall be required instead of a demonstration.</p>
<p>1. Copy of End User Acceptance Certificate from 1 government and 1 private institution</p>	<p>1. At least two (2) End User Acceptance Certificate preferably from 1 government and 1 private institution.</p>
<p>OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB:</p> <ul style="list-style-type: none"> • USB 1 should contain the Technical Components of the offering • USB 2 should contain the Financial Component of the offering 	<p>For omission</p>

F. IB2020-005F SUPPLY AND DELIVERY OF 1 UNIT COLPOSCOPE

FROM	TO
<ul style="list-style-type: none"> • Triseriescolposcopes 	<p>For omission</p>
<ul style="list-style-type: none"> • External USB cameras 	<p>Built-in camera or colposcope directly connected to a desktop for image view</p>
<ul style="list-style-type: none"> • Frequency range 4-12 MHz 	<p>For omission</p>
<p>2C. Service Unit:</p>	<p>2C. Service Unit: In case of Unit Malfunction, Bidder must provide a service unit within 48 hours until such time that the unit has been repaired or replaced.</p>
<p>3B. Preventive Maintenance</p>	<p>Bidder must conduct Quarterly Preventive Maintenance within the warranty period</p>
	<p>Bidder must have 24/7 readily available Technician to provide technical service and support within 24 hours from report via phone call or email in case of machine breakdown or as needs arise</p>
<p>– Should have end-user acceptance certificate from one government and one private institution</p>	<p>Must provide at least two (2) End User Acceptance Certificate preferably from 1</p>



Republic of the Philippines
 Department of Health, Center for Health Development (CHD) IV-CALABARZON
BATANGAS MEDICAL CENTER
 Batangas City
ISO 9001:2015 CERTIFIED



<p>OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB:</p> <ul style="list-style-type: none"> • USB 1 should contain the Technical Components of the offering • USB 2 should contain the Financial Component of the offering 	<p>government and 1 private institution.</p> <p>For omission</p>
---	--

G. IB2020-005G SUPPLY AND DELIVERY OF 1 UNIT ULTRASOUND MACHINE-OB

FROM	TO
3D/4D Ultrasound	With or without 5D Capacity
Lightweight and Maneuverable	For omission
High Resolution LCD, LED Display 22-25 inches	High Resolution LCD / LED Display at least 21inches
10-12 inches high resolution color LCD screen	10-12 inches high resolution color LED screen
Fully articulating monitor arm	For omission
4 active probe ports (Linear, Convex, Transvaginal, 4D probe)	4 active probe ports (Endovaginal 4D , Convex, Transvaginal, 4D probe)
Integrated HDD 500 GB – 1 Terabyte	Integrated HDD / SDD 500 GB – 1 Terabyte
Cross X Beam CRI TM (Compound Resolution Imaging)	Compound Resolution Imaging
Wide Sector (Max Angle)	Wide sector
12L-RS Wideband endocavity	For omission
Gel warmer with temperature control with three levels	External Gel warmer
2C. Service Unit:	2C. Service Unit: In case of Unit Malfunction, Bidder must provide a service unit within 48 hours until such time that the unit has been repaired or replaced.
	Bidder must have 24/7 readily available Technician to provide technical service and support within 24 hours from report via phone call or email in case of machine breakdown or as needs arise
3B. Preventive Maintenance	Bidder must conduct Quarterly Preventive Maintenance within the warranty period
– Should have end-user acceptance certificate from one government and one private institution	Must provide at least two (2) End User Acceptance Certificate preferably from 1 government and 1 private institution.
<p>OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB:</p> <ul style="list-style-type: none"> • USB 1 should contain the Technical Components of the offering • USB 2 should contain the Financial Component of the offering 	<p>For omission</p>

H. IB2020-005H SUPPLY AND DELIVERY OF 1 UNIT ELECTROCONVULSIVE THERAPY MACHINE

FROM	TO
<p>OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB:</p> <ul style="list-style-type: none"> • USB 1 should contain the Technical Components of the offering • USB 2 should contain the Financial Component of the offering 	<p>For omission</p>



I. IB2020-005I 1 UNIT ULTRASOUND MACHINE RADIOLOGY

FROM	TO
<ul style="list-style-type: none"> • 12.1 inch tilting multi-gesture touch screen 	<ul style="list-style-type: none"> • 12.1 inch tilting multi-gesture touch screen or its equivalent
<ul style="list-style-type: none"> • Gel Warmer with Temperature Control with three levels 	<ul style="list-style-type: none"> • Gel Warmer with Temperature Control with three levels (Integrated to the Machine / External)
<ul style="list-style-type: none"> • Built-in Wireless Adapter 	<ul style="list-style-type: none"> • Built-in Wireless Adapter (Optional)
<ul style="list-style-type: none"> • Echo Boost, HR Flow, iBeam, iClear • 3D iClear, Depth VR, MagiClean for volume image • Smart Doppler, Smart Track, iTouch, iZoom, V-Mapping 	<ul style="list-style-type: none"> • with improved visualization of tiny vessels and complex flow patterns • with enhanced B mode image profile for better boundary details • with spatial compound imaging, smoothness and edge enhancement adjustment for 2D and 3D • with more stereoscopic VR images • with automatic background and artifact cancelling function on volume probe • with automatic color box and doppler gate positioning • with vessel tracking and image optimization • with one key optimization function • with selective zoom function • with V-mapping or equivalent of each feature
<ul style="list-style-type: none"> • iWorks™ (Standardized workflow protocol) 	<ul style="list-style-type: none"> • Standardized workflow protocol
<ul style="list-style-type: none"> • iNeedle™ (Needle visualization enhancement) 	<ul style="list-style-type: none"> • Needle visualization enhancement or its equivalent
<p>Voice Command (audio control the system)</p>	<p>Voice Command (audio control the system) (Optional)</p>
<ul style="list-style-type: none"> • Biopsy kit: NGB-020, multi-angle, reusable 	<ul style="list-style-type: none"> • Biopsy kit: multiangle, reusable biopsy guide
<p>Endocavity Transducer</p> <ul style="list-style-type: none"> • FOV (max): 170° 	<p>Endocavity Transducer</p> <ul style="list-style-type: none"> • FOV (max): 170° or higher

<p>Training Foreign Training for 2 Consultants and 2 Radiologic Technologists with a Hands on Observations during Training</p>	<p>Training Virtual Training for Consultants, Residents, and Radiologic Technologists with Hands on Observations during in-house training. Apps training and refresher course shall be available upon request for at least 2 years.</p>
<p>5. Certification that the bidder will provide Trainings and certification for the following: (note: number of attendees will be provided by the end user and engineers)</p> <ol style="list-style-type: none"> a. At least 4-hours operation/end users training b. At least 4 hours of basic trouble shooting and maintenance for the engineers/technicians c. Recertification with in warranty as need by the end users and engineers/technicians 	<p>5. Certification that the bidder will provide Trainings and certification for the following: (note: number of attendees will be provided by the end user and engineers)</p> <ol style="list-style-type: none"> a. At least 4-hours operation/end users training b. At least 4 hours of basic trouble shooting and maintenance for the engineers/technicians c. Refresher course as needed by the end users within warranty period
	<p>Bidder must have 24/7 readily available Technician to provide technical service and support within 24 hours from report via phone call or email in case of machine breakdown or as needs arise</p>
<p>OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB:</p> <ul style="list-style-type: none"> • USB 1 should contain the Technical Components of the offering • USB 2 should contain the Financial Component of the offering 	<p>For omission</p>

J. IB2020-005J SUPPLY AND DELIVERY OF 1 UNIT MOBILE C-ARM SYSTEM

FROM	TO
<ul style="list-style-type: none"> • Compact mobile C-arm system that can be used in orthopedic surgery, vascular surgery, cardiac procedures, pain management procedures, abdominal procedures, thoracic surgery and general surgery 	<ul style="list-style-type: none"> • Compact mobile Flat Panel Detector C-arm system that can be used in orthopedic surgery, vascular surgery, cardiac procedures, pain management procedures, abdominal procedures, thoracic surgery, general surgery and urologic surgery
<ul style="list-style-type: none"> • Must have a radiation dose monitor/counter. 	<ul style="list-style-type: none"> • Must have a radiation dose monitor/counter integrated to the machine.
<ul style="list-style-type: none"> • The monitor cart should be at least 18 inches with minimum resolution of 1280x1024 	<ul style="list-style-type: none"> • The monitor screen size should be at least 18 inches with minimum resolution of 1280x1024
<ul style="list-style-type: none"> • Must have a video paper printer 	<ul style="list-style-type: none"> • Must have a thermal paper printer
<ul style="list-style-type: none"> • Overscan of C-arm: at least 100 degrees 	<ul style="list-style-type: none"> • Overscan or its equivalent of C-arm: at least 100 degrees
<ul style="list-style-type: none"> • Must provide compatible sterile surgical 	<ul style="list-style-type: none"> • Must provide and can supply compatible

	Republic of the Philippines Department of Health, Center for Health Development (CHD) IV-CALABARZON BATANGAS MEDICAL CENTER Batangas City ISO 9001:2015 CERTIFIED	
--	---	--

drape/covers.	sterile surgical drape/covers Bidder must have 24/7 readily available Technician to provide technical service and support within 24 hours from report via phone call or email in case of machine breakdown or as needs arise
• Must provide a copy of end user acceptance certificate from 1 government and 1 private institution.	Must provide at least two (2) End User Acceptance Certificate preferably from 1 government and 1 private institution.
OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB: <ul style="list-style-type: none"> • USB 1 should contain the Technical Components of the offering • USB 2 should contain the Financial Component of the offering 	For omission

K. IB2020-005K SUPPLY AND DELIVERY OF 1 UNIT STANDARD PERCUTANEOUS NEPHROLITHOTOMY

FROM	TO
1 PC Otis Urethrotome length 16cm with 2 knives	1 PC Otis Urethrotome length 16cm with 2 knives (Optional)
1 PC Lithotrite 24Fr	1 PC Lithotrite 24Fr / Lithotripter
	Bidder must have 24/7 readily available Technician to provide technical service and support within 24 hours from report via phone call or email in case of machine breakdown or as needs arise
	In case Bidder failed to deliver the Unit within the delivery schedule, Bidder must provide a service until such time that the unit has been delivered.
OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB: <ul style="list-style-type: none"> • USB 1 should contain the Technical Components of the offering • USB 2 should contain the Financial Component of the offering 	For omission

L. IB2020-005L 1 UNIT OPHTHALMIC OPERATING MICROSCOPE 90 w/ X-

FROM	TO
Focus Range:48mm - 50mm	Focus Range: 48mm - 54mm
Magnification: Motorized zoom (zoom ratio 1:6) approx	Magnification: Motorized zoom / Motorized magnification changer (zoom ratio 1:6) approx
Binocular Tube: 45° binocular inclined with converging optics/180° - 200° swivel tube (swivel tube preferred)	Binocular Tube: at least 45° binocular inclined with converging optics/180° - 200° swivel tube (swivel tube preferred)
Total Magnification range: 4.28x to 21.4x	Total Magnification range: 3.4x to 30x





Republic of the Philippines
 Department of Health, Center for Health Development (CHD) IV-CALABARZON
BATANGAS MEDICAL CENTER
 Batangas City
ISO 9001:2015 CERTIFIED



Filters: Heat absorbing, UV filter, Yellow filter, Cobalt blue, Red Free	Filters: At least with yellow filter/halogen mode and blue light filter
Arm Vertical Stroke: Approx ≥ 400 mm	Arm Vertical Stroke: Approx ≥ 300mm
ACCESSORIES INCLUDED • Camera attachment • Beam Splitter	ACCESSORIES INCLUDED • Camera attachment(optional) • Beam Splitter(external / integrated)
Warranty period 1 – 2 years	Warranty period 2 years
	For demonstration/confirmation of specification during the evaluation
OTHER TERMS AND CONDITIONS 1. Copy of End User Acceptance Certificate from 1 government and 1 private institution. 2. Training should be once a year for 2 years. 3. Preventive and corrective maintenance scheduled quarterly.	OTHER TERMS AND CONDITIONS 1. At least two (2) End User Acceptance Certificate preferably from 1 government and 1 private institution. 2. Training should be once a year for 2 years. 3. Preventive and corrective maintenance scheduled semi-annually for two (2) years.
	Service Unit: In case of Unit Malfunction, Bidder must provide a service unit within 48 hours until such time that the unit has been repaired or replaced.
	Bidder must have 24/7 readily available Technician to provide technical service and support within 24 hours from report via phone call or email in case of machine breakdown or as needs arise
OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB: • USB 1 should contain the Technical Components of the offering • USB 2 should contain the Financial Component of the offering	For omission

M. IB2020-005M SUPPLY AND DELIVERY OF 1 UNIT DIGITAL FLUOROSCOPY MACHINE

FROM	TO
FLUOROSCOPY TABLE Dimensions 225x159x293 max horizontal or better	FLUOROSCOPY TABLE Dimensions 225x81 cm. or better with horizontal coverage of 293 cm.
Weight Capacity 230kgs or better	Weight Capacity At least 180 KG at any point in the table
Digital Portable/Mobile X-Ray (Motorized) X-RAY TUBE 0.7/1.3 Focal Spot	Digital Portable/Mobile X-Ray (Motorized) X-RAY TUBE At least 0.6/1.3 Focal Spot
POSITIONING +/- 315 Degrees Column Rotation or better	POSITIONING At least +/-270 Degrees Column Rotation
UPS	2 UPS One (1) for the Fluoro console and One (1)

	Republic of the Philippines Department of Health, Center for Health Development (CHD) IV-CALABARZON BATANGAS MEDICAL CENTER Batangas City ISO 9001:2015 CERTIFIED	
--	---	--

	for the Radiologist workstation
	Provision of Bucky wall
C. Service Unit:	For omission
	Bidder must provide Apps Training / Refresher Course to the end-users as needed for at least 2 years
	Bidder must have 24/7 readily available Technician to provide technical service and support within 24 hours from report via phone call or email in case of machine breakdown or as needs arise
OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB: <ul style="list-style-type: none"> • USB 1 should contain the Technical Components of the offering • USB 2 should contain the Financial Component of the offering 	For omission

N. IB2020-005N SUPPLY AND DELIVERY OF 1 UNIT STEAM STERILIZER

FROM	TO
POWER: 48KW	POWER: 45-50KW
GENERATOR CAPACITY 55 LITERS.	GENERATOR CAPACITY 50-80LITERS
CHAMBER DIMENSIONS (mm) : 670 x 670 x 998 (HEIGHT X WIDTH X DEPTH)	CHAMBER DIMENSIONS (mm): HEIGHT: 650-750 WIDTH: 650-750 DEPTH 900-1000
OVERALL DIMENSIONS (mm) : 1954 x 996 x 1314 (HEIGHT X WIDTH X DEPTH)	OVERALL DIMENSIONS (mm) : HEIGHT: 1900-2500 WIDTH: 900-1000 DEPTH: 1250-1350
CHAMBER AND STEAM MANOMETERS AT THE LOADING SIDE AND CHAMBER MANOMETER AT THE UNLOADING SIDE.	CHAMBER AND STEAM MANOMETERS OR LCD DISPLAY MANOMETER AT THE LOADING SIDE AND CHAMBER MANOMETER AT THE UNLOADING SIDE.
OTHER REQUIREMENTS TO BE SUBMITTED ON THE BID OPENING: Bidder must provide two (2) USB: <ul style="list-style-type: none"> • USB 1 should contain the Technical Components of the offering • USB 2 should contain the Financial Component of the offering 	For omission

III. Amendment of Schedule of Requirements as per the Pre-Bidding Conference conducted last August 22, 2020 are as follows:

A. IB 2020-005B B SUPPLY AND DELIVERY OF OR RHINOLOGY SET

FROM	TO
60-90 DAYS FROM THE RECEIPT OF PURCHASE ORDER	60-120 DAYS FROM THE RECEIPT OF PURCHASE ORDER



Republic of the Philippines
 Department of Health, Center for Health Development (CHD) IV-CALABARZON
BATANGAS MEDICAL CENTER
 Batangas City
ISO 9001:2015 CERTIFIED



B. IB 2020-005C SUPPLY AND DELIVERY OF OR LARYNGOSCOPE SET

FROM	TO
60-90 DAYS FROM THE RECEIPT OF PURCHASE ORDER	60-120 DAYS FROM THE RECEIPT OF PURCHASE ORDER

C. IB 2020-005D SUPPLY AND DELIVERY OF BRONCHOSCOPE SET (ADULT AND PEDIA)

FROM	TO
60-90 DAYS FROM THE RECEIPT OF PURCHASE ORDER	60-120 DAYS FROM THE RECEIPT OF PURCHASE ORDER

D. IB2020-005K SUPPLY AND DELIVERY OF 1 UNIT STANDARD PERCUTANEOUS NEPHROLITHOTOMY

FROM	TO
60-90 DAYS FROM THE RECEIPT OF PURCHASE ORDER	90-120 DAYS FROM THE RECEIPT OF PURCHASE ORDER

ELIZABETH V. PALINES, MD, FPPS, FCNSP, FPNA
 Chairperson, HBAC

Received by the bidder:

 Signature over printed name
 Date Received: _____